SGS

EC Certificate Production Full Quality Assurance System: Certificate ES19/86904

The management system of

MEDICAL PRECISION IMPLANTS, S.A.

C/ Empleo, 21, P.I. Los Olivos, 28906 Getafe, Madrid Spain

has been assessed and certified as meeting the requirements of

Directive 93/42/EEC

on medical devices, Annex V

For the following products

- Non sterile dental drills.
- Non sterile dental taps.

Class I Measuring: "Metrological aspects only - Restricted to the aspects of manufacture concerned with the conformity of the devices with metrological requirements":

- Non-sterile dental depth gauges

Where the above scope includes class IIb or class III medical device(s), a valid EC Type Examination Certificate according to Annex III is a mandatory requirement for each device in addition to this certificate to place that device on the market.

This certificate is valid from 20 April 2021 until 24 May 2024 and remains valid subject to satisfactory surveillance audits.

Issue 4. Certified since 11 February 2011

Certification is based on reports numbered ES/MAD/ 148832

Authorised by

Global Medical Devices Head of Notified Body

asselvi

SGS Belgium NV, Notified Body 1639

SGS House Noorderlaan 87 2030 Antwerp Belgium t +32 (0)3 545-48-48 f +32 (0)3 545-48-49 www.sgs.com

LPMD5008 - Certificate CE1639 AnnexV_EN Rev. 02

Page 1 of 1





This document is issued by the Company subject to its General Conditions of Certification Services, unless otherwise agreed, accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability, indermification and jurisdictional issues established therein. The authenticity of this document may be verified at https://www.sgs.com/en/certified-clients-and-products/certified-client-directory. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.